

# AYO<sup>TM</sup> HC

## RESPIRATOR



USER MANUAL

PRODUCT CERTIFICATION



BSI Certified Product  
AS/NZS 1716:2012 Lic: BMP No 793811

## The World's First Healthcare Reusable Respirator

## CONTENTS

Foreword	3
Introduction	3
Warnings and Safety Instruction	3
Know Your AYO HC	4
Product Configuration	4
AYO HC Components	5
How to Use	6
Module Connection	6
Donning	7
Doffing	8
Module Disconnection	8
Changing the filters	9
How to Select the Mask Size	11
Quick Fit Check	11
Fit Testing	12
Maintenance	17
Storage	19
Replacement / Consumable Items	20
Technical Specifications	21
Troubleshooting	22
Other information	22
Disclaimer	23
Limited Liability	23
Warranty	23
Contact	23

## Foreword

Before using the AYO HC Air-Purifying Respirator (APR) users must read and understand the User Manual. Keep these instructions for reference. Failure to follow all instructions for the use of this product may lead to adverse effects on the users' health and may render the warranty void.

## Introduction

Congratulations on purchasing the AYO HC Air Purifying Respirator, the world's first healthcare reusable respirator.

This certified product provides more reliable and enhanced protection than traditional N95 masks. AYO HC sets a new standard in healthcare and non-industrial respiratory protection.

## Warnings and Safety Instruction

### Safety Instructions

AYO HC helps protect against infectious aerosol transmission, pollen, city air pollution, and general particulate contaminants in the air if it is used correctly. The following safety instructions must be followed to avoid possible harm to your safety and health:

- Inspect the respirator before each use. Do not use the mask if it is damaged or has missing parts.
- Filters need to be regularly changed. The details are explained in Chapter 5 (How to use - Changing Filters).
- Store the respirator in a clean and dry area when not in use.

### Warnings and Limitations

- Do not use the respirator in atmospheres that are dangerous to life or where the identity and concentration of contamination are unknown.
- Do not use the respirator for escape purposes.
- Do not remove the respirator unless you have left the contamination area.
- Do not substitute, modify, add, or omit any components of this product.
- Use by clean-shaven personnel only. Facial hair under a seal reduces protection.

## ■ Know Your AYO HC

### Product Configuration

The AYO HC comprises three major components: **Mask Assembly (MA)**, **Outlet Module (HCo)** and **Inlet Module (HCi)**.





Figure 1 - AYO HC Configuration

## AYO HC Components

The AYO HC is designed to filter air intake whilst being worn. Detachable HCo (2) and HCo (3) are fitted to an AYO MA mask assembly (1) which is attached to the face with an adjustable head strap (4).




The AYO MA Mask Assembly comprises a silicone mask, a bridge assembly, and a removable head strap.

The bridge assembly consists of a speech diaphragm (5) to facilitate speech and an inlet non-return valve (6) to direct airflow during use.

Adjustment of the head strap tension combined with a proper silicone mask selection (Small, Medium, Large) will ensure good comfort and air seal.

## How To Use

### Module Connection

 It is important for the modules to be connected to the correct side of the mask.

The Inlet module HCo (3) is to be fitted to the left side of the AYO MA (1), and the Outlet Module HCo (2) to the right side of the AYO MA (1).




### Connecting Modules to Mask

Connection of modules is achieved by pushing the Push-Fit Plugs on the mask (4) to the socket of HCo /HCo (5) until a **'Click'** sound is heard and the AYO MA is reliably connected with the modules.




## Donning

- Press the strap lock button (6) to adjust the strap length/tension and establish a roughly suitable strap length.
- Place the strap on the back of the neck and place the mask on the face.
- Slide two head pads (7) to the top and the back of the head in the position shown.
- Adjust strap tension for comfort and seal using the strap lock button (6).
- Adjust the position of the head pads by sliding the adjusting sleeve (8) so that the head pads are in stable positions on the head as shown in the figure below.




Ensure that side straps are threaded through the hook rings on both modules.




## Doffing

- Pull the two pads off the head and take the back strap off from the neck. (Note: You may keep the strap lock button position unchanged for the next use to avoid adjustment every time.)
- Remove the mask from the head.



## Module Disconnection

To disconnect the modules from the mask, simply turn the grip (1) while holding the module case (2) to release the module.






## Changing the filters

### Changing the Main Filter in Inlet Module - HCl

The HCl module contains a cartridge filter that removes particulate from inhaled air.

**Step 1:** Press Module button (7) and remove End Cap (8) out – the filter cartridge (9) is now ready for replacement.



**Step 2:** Replace Filter Cartridge

When inserting the filter, align the markings of the filter cartridge and module housing for correct orientation.

#### Note:

- Change the filter when it is contaminated, its breathing resistance increases, or it has been used for over three months.
- Shelf Life: Maximum of 5 years from the year of manufacture for unopened filters stored at room temperature and 50% relative humidity.
- Only genuine AYO HC filter cartridges can be used.

**Warning:** Filter performance is affected by isopropyl alcohol (IPA). Do not use with isopropyl alcohol to ensure protection.

## Changing the Exhalation Filter in Outlet Module - HCo

The HCo module contains an optional exhalation filter to remove droplets from exhaled air. The procedure for replacement is as follows:

### Optional Cartridge filter

**Step 1:** Press the Module button (7) then remove End Cap (8) and filter Cartridge Holder (10) from the HCo module housing (12).

**Step 2:** Remove the filter cartridge (11)



**Step 3:** Fit a new cartridge (11) into the cartridge holder (10) taking care that the cartridge holder is aligned correctly with End Cap (8)



**Step 4:** Insert the cartridge holder (10) into housing (12) then close with the end cap (8).



Note: Ensure that the exhalation seal and the round feature on the end cap are aligned before assembling the end cap.

Warning: pay attention not to accidentally disturb the Exhalation Membrane, and ensure it sits flat on its seat.


## How To Select the Mask Size

In most cases, mask size can be selected based on the length from the top of the nose bridge to the dent between the lower lip and the chin – the Z dimension:

- Small Size: from 68 mm to 80 mm
- Medium Size: from 80 mm to 100 mm
- Large Size: from 100 mm to 113 mm

To ensure a good fit, double-check with the actual mask to see:

- If it fits your face snugly.
- If the mask cushion sits on top of your nose bridge and around the midpoint between the bottom lip and the chin.



### Quick Fit Check

1. Don the mask assembly (MA) on your face without fitting the Inlet and Outlet modules.
2. Block each end of the MA with your thumb.



3. Try to inhale. If mask is pressed against your face by your inhaling and there is no signs of facial leak around the cushion of the mask, the fit is satisfactory.

If you feel or hear air leaks around the cushion (especially nasal bridge area), it may indicate an incorrect fit, and you may try a different mask size and repeat the above check until the fit is satisfactory.

## ■ Fit Testing

To check the mask-fit thoroughly, we recommend conducting a proper Fit Testing for the first time and subsequently every 12 months thereafter with specialized Fit Testing equipment, such as 'PortaCount Respirator Fit Tester 4048'. Fit Test can be provided by suitable third-party services which are widely available.

### Fit Testing Set-up Procedure

#### Step 1

##### **Pass Tube Holder (4) through the Exhalation side of the Mask (5).**

The HC Fit Test Adaptor is composed of Adaptor Main Body (1), Fitting Port (2), Silicone Tube (3) and Tube Holder (4).

Pass Tube Holder (4) with the silicone tube (3) through the Exhalation side of the Mask (5).



**Fit Tube Holder.**

Simply press Tube Holder (4) onto the adjacent slots of Bridge Assembly.

It is recommended to install this Tube Holder (4) near the right side, so that the probing point is close to the centreline of the lips.



4

## Step 2

### Connect Fit Test Adaptor

Connect Fit Test Adaptor socket side (6) to the Push-Fit Plugs in the exhalation side of Mask (5).


Fitting port (2) is recommended to face down for the convenience of connecting a tube from Fit Test Machine.



### Step 3

#### Connect the HCo Unit.

Connect a Fit Test Adaptor plug side (7) to the connecting socket of the HCo Unit (8).



### Step 4

#### Connect HCo (9)

Change the Filter in HCo from AP056 to AP096 higher efficiency filter, then connect HCo (9) to Mask (10)

Note that hooks of HCo (11) and HCo (11) should face inside to ensure correct flow direction of HCo.



## Step 5

### Connect Fit Testing Tube

Connect Fit Testing Tube from Fit Test Machine to the Fitting Port (2) on the Fit Test Adaptor.

Now the AYO HC is ready for Fit Testing. To know the remaining steps and details for conducting Fit Testing, refer to the instructions of the Fit Tester being used (e.g. TSI PortaCount Respirator Fit Tester).



## Step 6

### Remove the Fit Test Adaptor

When remove the Adaptor, do exactly reverse steps as the installation procedure:

- 1) Remove the Tube Holder from Bridge Assembly.
- 2) Disconnect the Fit Test Adaptor from the mask and from HCo and slide out the Tube and Tube holder from exhalation side of the mask.

After removing Fit Test Adaptor, connect HCo for normal use.





---

## ■ Maintenance

### **Routine Visual examination:**

- Replace AYO Mask if there are any signs of hardening, signs of stress, cracks, tears and deformation.
- Replace Head Straps if there are any signs of wear.
- Replace Filter if there is any damage or cracks.
- Replace HC case if there is any damage or connection issue.

### **Daily cleaning:**

After each long use,

- a) Detach H<sub>Ci</sub> and H<sub>Co</sub> from AYO MA
- b) Cleaning of AYO MA: Apply tap water plus a few drops of suitable detergent such as mild dishwashing liquid or laundry liquid to handwash inside the mask including the silicone mask, bridge piece, and inlet membrane - paying attention to skin contact surfaces. Flush wash the whole AYO MA inside and outside.
- c) Use a dry cloth to dry out the mask from inside to outside, then leave it in a well-ventilated clean area until drying completely.
- d) Drying of H<sub>Co</sub>: remove the End Cap of the H<sub>Co</sub>, then take out the Cartridge Holder and leave all the disconnected parts in a well-ventilated clean area until drying completely.

## Thorough cleaning

For frequent use of AYO HC, it is advised to conduct a thorough cleaning each month.

a) Detach the H<sub>Ci</sub> and H<sub>Co</sub> from the AYO MA, then take off the head strap from the mask by pressing the strap lock button and pulling the head strap out of the lock button.

After that, partially peel off the silicone mask gently from the bridge assembly while keeping the front cover unremoved.


- b) Remove the EndCap on both H<sub>Ci</sub> and H<sub>Co</sub>. Remove the main filter and the optional exhalation filter.
- c) Soak all parts in soapy water or suitable detergents such as mild dishwashing liquid or laundry liquid for 10 minutes, except the main filter element and the exhalation filter.
- d) Thoroughly hand-wash inside and outside of the AYO MA, especially the area that you cannot get access to during daily cleaning, then flush wash all the parts inside and outside.
- e) Drain and shake off water from all the parts. Then air dry or use a clean and dry cloth to rub dry further.
- f) Leave all the washed parts in a well-ventilated clean area drying completely.
- g) Carefully fit the Bridge Assembly back into the mask and ensure the end of the Bridge Assembly fits properly inside the mask.



Make sure the edge of Bridge Assembly aligns with the locating rims inside the mask.



h) Finally, make sure the edge of the bottom tab of the Bridge Assembly is fitted in the latching groove at the bottom of the mask.



i) Assemble the H<sub>Ci</sub> and H<sub>Co</sub> modules back for next use, and make sure the membrane in the Exhalation Filter Cartridge Holder sits nice and flat in its seat.

## ■ Storage

- Daily Storage: store AYO HC components in the casing as provided or in a suitable, clean environment.
- When not in use for an extended time, ensure the casing is properly covered and protected from direct sunlight, dust, extreme temperature, excessive moisture, and harsh chemicals. (Please refer to 'Technical Specifications' for storage conditions)

For more information about maintenance and storage of AYO HC, refer to **AS/NZS 1715**.

## ■ Replacement / Consumable Items


The table below shows replacement/consumable parts when needed.

The replacement/consumable parts are available online:

<https://aimwellbreathing.com/shop-healthcare-respirator/>

Item name	Item Code
AYO HC Medium	AP089
AYO HC Small	AP090
AYO HC Large	AP091
AYO MA (Mask Assembly)	Part No. AP050Q (Medium)
	Part No. AP051Q (Small)
	Part No. AP052Q (Large)
AYO HC Inlet Filter Element 95 (Pack of 3)	AP056
AYO HC Outlet Filter Element - Droplets (Pack of 3)	AP092
AYO HC Head Strap Assembly	AP055
AYO HC Inlet Module - HCl	AP054i
AYO HC Outlet Module - HCo	AP054o
AYO HC Bridge assembly (Speech Diaphragm and membrane)	AP084
AYO HC Mask Speech Diaphragm Cover	AP085
AYO HC Fit Testing Kit	AP086

## ■ Technical Specifications

<b>Operating conditions</b>	
Operating Temperature Range	-20°C to 45°C
Operating Relative Humidity Range	Up to 90% RH non-condensing
<b>Storage Conditions</b>	
Temperature Range	-20°C to 70°C
Shelf life for Filters	Maximum of 5 years from the year of manufacture for unopened filters stored at room temperature and 50% relative humidity.
<b>Dimension and Weight (Exhalation and Filter modules)</b>	
Dimensions	103 mm x 43mm x 57mm
AYO HC Outlet Module – HCo Weight	Up to 80g
AYO HC Inlet Module - HCo Weight	Up to 75g
<b>Standards Compliance</b>	
Australia	AS/NZS 1716:2012 P2
China	GB 2626 KN95 (TBD)
EU	EN140 / EN143 P2 (TBD)
Certified by: BSI Group ANZ Pty Ltd, Suite 1, Level 1, 54 Waterloo Road, Macquarie Park, NSW 2113 Australia	 <p>PRODUCT CERTIFICATION</p> <p>BSI Certified Product</p> <p>AS/NZS 1716:2012 LK:BMP No 793811</p>

## ■ Troubleshooting

Problem	Cause and remedy
The mask plug comes off from the attached module by itself.	<p>When connecting the mask with any of the modules, always push sufficiently until the mask plug is fully latched in the connection socket of the module with a <b>'Click'</b> sound.</p> <p>Make sure the mask plug sits fully in the connecting socket and cannot be pulled out of the socket.</p>

## ■ Other Information

### Key Materials

Mask: Medical-grade silicone rubber

Casing: polycarbonate

Head strap: nylon, neoprene, and synthetic leather

### Further Resource

More information related to AYO HC can be found via the website: <https://aimwellbreathing.com>

---

## ■ Disclaimer

You acknowledge and agree that at all times you will use AYO HC in accordance with any user documentation, instruction manuals, guides, and/or requirements we provide to you.

Aimwell makes no representations or warranties and shall not be responsible for the competency of any person who uses the AYO HC. If you do not feel well using the AYO HC or have difficulty in breathing, please stop using the AYO HC.

## ■ Limited Liability

To the extent permitted by law, you will not hold Aimwell liable for any personal injury, loss of use, loss of profit, interruption of business, or any indirect, special, incidental, punitive, or consequential damages of any kind (including lost profits), regardless of the form of action, whether in contract, tort (including negligence), strict product liability, or otherwise, even if Aimwell has been advised or should have been aware of the possibility of such damages. To the extent permitted by law, Aimwell's liability to you for any claim, whether in contract, tort, or any other theory of liability, does not exceed the purchase price of the system paid by you.

It is further understood that Aimwell shall also have no liability whatsoever concerning damages arising out of or relating to unauthorized access, use, and/or alteration or destruction, by any third party, including any users, of the system or any data contained thereon or created by the system or your computer system.

## ■ Warranty

The warranty is valid for 12 months from the date of purchase, during which we offer repair or replacement of faulty products free of charge.

## ■ Contact

info@aimwellbreathing.com  
(02) 7251 8689



PRODUCT CERTIFICATION



BSI Certified Product  
AS/NZS 1716:2012 Lic:BMP No 793811



scan to visit the site  
[www.aimwellbreathing.com](http://www.aimwellbreathing.com)



[info@aimwellbreathing.com](mailto:info@aimwellbreathing.com)



(02) 7251 8689